

**ACS Change Log**  
**(Newest Version at bottom of list)**

**4/26/10 – Version 4.5.0 (100,200,600) full installers on FTP**

Fixed crash on Tymp Compliance test - stack overflow on GotoSignal  
Removed spaces in the customer and audiometer info (dao-null issue)

**5/5/10 – Version 4.5.0.1 (100,200,600) full installers on FTP**

Fixed crash when converting v3.0 or v3.5 files  
Compiled Titan\_IO.dll in release mode  
Changed SIB parameter to A3x to H1x for HF phones  
Reenabled Earscan audiometers in model list

**5/18/10 – Version 4.5.0.2 (100,200,600) full installers on FTP**

Titan\_IO.dll and Interacoustics DLL's built in release mode

**6/02/10 – Version 4.5.0.3 (100,200,600) full installers on FTP**

Fixed White Noise Total Power Cal on GSI-61

**6/04/10 – Version 4.5.0.4 (100,200,600) full installers on FTP**

Titan changes:  
Added 1300 and 2000 Hz for Probe tests  
Added Cavity Calibration logic (changed Titan\_IO.dll)  
Disabled Reflex Switching in default cfg  
Cal level parameter now using SPL instead of HL  
Fixed HF Narrowband with no 1/3 octave filtering

**6/25/10 – Version 4.5.0.5 (100,200,600) full installers on FTP**

Revised Titan Tymp Compliance calibration logic

**6/30/10 – Version 4.5.0.6 (100,200,600) full installers on FTP**

Fixed bug importing HF settings from older config file

**7/11/10 – Version 4.5.0.7 (100,200,600) full installers on FTP**

Logic change to always save cal on GSI-61 regardless of adjustment threshold

**7/19/10 – Version 4.5.0.8 (100,200,600) full installers on FTP**

Changed GSI-61 bone maxHL for 8K from 80 to 75

**7/20/10 – Version 4.5.0.9 (100,200,600) full installers on FTP**

Fixed bug not allowing 12K frequency selection

**8/4/10 – Version 4.5.0.10 (100,200,600) full installers on FTP**

Completed Titan Compliance Calibration Sequence  
Rev Titan\_IO.dll v1.0.0.2

**8/10/10 – Version 4.5.0.11 (100,200,600) full installers on FTP**

Added option to edit user info before each cal  
Added support for Welch Allyn AM282 (GSI-18)  
Added support for Welch Allyn TM286 (GSI-39)  
Preliminary (incomplete) SIB PRIMUS logging

**8/14/10 – Version 4.5.0.12 (100,200,600) full installers on FTP**

Completed SIB PRIMUS logging file  
Added support for Belltone Audio Scout  
Added support for Belltone 119

**8/18/10 – Version 4.5.0.13 (100,200,600) full installers on FTP**

Added Transducer SN Prompt for SIB PRIMUS logging file

**8/23/10 – Version 4.5.0.14 released**

SIB Log file changes:

- Changed log file cal value from dB10 to dB
- Actual log filename written to first line of log file
- Changed text prompt on log file S/N to "Enter the original transducer manufacturer serial numbers:"
- The original transducer S/N's and the 5 digit code written to log file
- Left and Right modifiers added to coupler information in log file
- Improved the logic for validating microphone equipment setup before calibrating
- Added a "Log File Path" field to the Serial Setup Dialog box.

**9/3/10 – Version 4.5.0.15 released**

Added Interacoustics AS608 (manual only for now)  
Optimized DT\_io (keep A/D subsystem open until Close(), free low level DT buffers on Close())  
Optimized recovery from DT buffer overrun  
Reformatted some of the screens to fit on a 10" netbook screen (buttons now on top right)  
Used new RETSPL for 12.5K for HDA-280

**9/28/10 – Version 4.5.0.16 released**

Fix for DT crash on Win7 (subsystem handle scope)

**9/30/10 – Version 4.5.0.17 released**

GUI improvements for Hardware Certification version for Interacoustics

**11/5/10 – Version 4.5.0.18 released**

Added Interacoustics DD45 headphones

**11/5/10 – Version 4.5.0.19 released**

Updated RETSPLs for DD45 for both coupler types  
Set Impedance to 10 Ohms if DD45 selected  
GSI-18 and AM282 – reduced MaxHL at 6kHz and

**12/1/10 – Version 4.5.0.20 released**

Changed transducer name from “Interacoustics DD45” to “DD45”  
Corrected log filename label for HDA280 for SIB

**12/2/10 – Version 4.5.0.21 released**

GSI-18/AM282 added button to “Reset Default Cal Data” on AudInfo

**12/20/10 – Version 4.5.0.22 released**

GSI-18/AM282 “Reset Default Cal Data” – removed check for ACK

**12/21/10 – Version 4.5.0.23 released**

Fixes for storing transducer info on eeprom for SIB Primus

**1/20/11 – Version 4.5.0.24 released**

Affinity/Equinox/Unity2 – Thor changes

**1/24/11 – Version 4.5.0.25 private release**

Titan Compliance – cal all points, exclude during validation only  
Titan\_IO.dll – added call to CalibrateGain() after probe cal

**2/15/11 – Version 4.5.0.26 released**

Support for Maico EasyTymp  
Support for Interacoustics Affinity  
Support for Interacoustics Equinox  
Siemens Unity2 interface now common with Affinity/Equinox  
Support for ANSI S3.6 2010  
Added a Message box warning if ACS System Calibration date is near, exceeded, or formatted incorrectly  
Added AUSSCO.INI hook [Preferences] “DisableCustomerControls” to allow you to disable customer controls on main screen. To use set DisableCustomerControls=1

**3/10/11 – Version 4.5.0.27 released**

Changed USBCommunication.dll to release version in installer  
Rebuilt Thor\_IO.dll to add version info

### **3/11/11 – Version 4.5.0.28 released**

Fixed Affinity/Equinox loss of signal problem with insert phones (Thor/Titan DLL's v1.0.0.4).  
Added a retest limit on the Rise and Fall test sections of Electrical Test  
Calibration data is automatically saved to "Recovery Data.mdb" file after each test closes to allow recovery of cal data if system crashes

### **3/19/11 – Version 4.5.0.29 released**

Thor Changes:

Added Free Field output selection to Transducer setup dialog  
Changed BC White noise Retspl to 42.5 (manufacturer spec)

Other:

Added logic to fix problem causing Bone to have only 1kHz selected  
Fixed problem with certificate layout that was using extra pages  
Fixed bug when re-editing audiometer info for user defined tympanometry

### **4/27/11 – Version 4.5.0.30 released**

Thor – Free Field calibration implemented (Thor\_IO, Titan\_IO v1.0.0.7)

### **4/28/11 – Version 4.5.0.31 released**

Private test versions:

4.5.0.32 – Maico MA-33 test  
4.5.0.33 – Callisto Test  
4.5.0.34 – Maico MA-33 with attenuator cal (Callisto disabled)  
4.5.0.35 – Welch Allyn Audioscope (Callisto disabled)  
4.5.0.36 - Maico MA-33 –init library once (Callisto disabled)  
4.5.0.37 – Rise/Fall/Over spectrum algorithm (Callisto disabled)  
4.5.0.38 – Callisto private release – save transducer setup fix

v4538 – added support for AMC493-B humidity tables.

Maico MA-40, MA-41, MA-42

Maico MA-33 (manual for now)

Restrict freq span to 3% when measuring electrical ref levels

Added White Noise upper bandwidth switch to User Prefs.

Log Mic Cal Data to .mic/.mst file

### **8/29/11 – version 4.5.0.39 released**

Added Interacoustics AD629,

Added Interacoustics Callisto

Added Interacoustics AT235, AT235H

Added support for Interacoustics AA222 Tympanometry calibration

Optimized communication logic for AA222/AT235

Fixed Insert Masking channel routing for Thor, AD629  
Fixed HF Narrowband cal on AD629 and Callisto  
Added Speech/Speech Noise Equivalent FF  
Added Speech/Speech Noise Non Linear  
Added Speech Noise ILTASS  
Added Speech Noise TENoise  
Added Reflex Speech Noise and White Noise  
Updated RETSPL's for HF Phones for 18K and 20K per Interacoustics docs  
Updated RETSPL's for DD45 per latest Interacoustics docs  
Fixed bug in electrical test requiring a "Continue" press when only Insert and HF phones used  
Fixed COM4 limit problem with Benson audiometers  
Fixed bug when running Linearity acoustically with HF phones

Updated Interacoustics DLL's for AD629, Callisto, Thor, Titan

AD629_IO.DLL	v1.0.0.2
Callisto_IO.DLL	v1.0.0.4
Thor_IO.DLL	v1.0.0.9
Titan_IO.DLL	v1.0.0.7
BensonControl32.DLL	v1.2.0.0 (allows ports > COM4 to be used)

#### **12/12/11 – version 4.6.0.0 released**

Added support for Holmco 8103 on MA-33  
Corrected signal setup and routing when insert masking phone used on Ad629 and Thor  
Reduced default test HL for certain HF test points that were above maxHL on Ad629 and Thor  
Reduced default test HL for insert phones that were above maxHL for MA-33  
CIR22 - eliminated unnecessary channel prompting during fully automatic testing  
Clarified error messages with Thor/Ad629/Titan

Updated:

AD629_IO.DLL	v1.0.0.3
Thor_IO.DLL	v1.0.0.10

#### **12/23/12 – version 4.6.0.1 released**

Fixed crosstalk input channel indexing error when running electrical tests on Inserts or HF Phones on ACS-600  
Thor\_IO – fixed transducer logic error when both TDH and HF phones selected.

#### **1/12/12 – version 4.6.0.2 released (also Thor\_IO.DLL 1.0.0.11)**

Thor – fixed transducer setup for HF/HLS phones.

#### **2/21/12 - version 4.6.0.3 released (also Thor\_IO.dll 1.0.0.12)**

MA-33 serial IO enabled  
Installer built with MA-33 DLL set instead of Thor/Titan  
Fixed EquivFF/NonLinear right channel missing values on certificate  
Amplivox 116 – manual only

**3/3/12 - version 4.6.0.4 released (MA-33 support but no Thor/Titan)**

RETSPLs updated for Koss HF phones at 18k and 20k  
TENoise RETSPL set to 25.0 (all transducers and specs)

**3/21/12 - version 4.6.0.5 released (No MA-33 support)**

Sennheiser HDA-280 added to headphone list  
Thor/Ad629/Callisto transducer setup logic for HDA-280 changed  
New IA Calibration Pack (received Feb 15 2012) installed. DLL's recompiled:  
Thor\_IO.dll v1.0.0.13  
Callisto\_IO.dll v1.0.0.5  
AD629\_IO.dll v1.0.0.4  
Titan\_IO.dll v1.0.0.8

**3/23/12 - version 4.6.0.6 released (No MA-33 support)**

**3/23/12 - version 4.6.0.7 released (MA-33 support but no Thor/Titan)**

MA-33 changes:  
SetTransducer logic  
Warble test  
Speech Level test tone – added delay for setup latency

**4/24/12 – v4.6.0.8 test exe - private**

Otovation Otosphere interface added

**4/26/12 – Version 4.6.0.9 to Otovation**

Otosphere Transducer Table utility added

**5/8/12 – Version 4.6.0.10 to Otovation**

Enabled Ampivox control logic for testing

**5/15/12 – Version 4.6.0.11 – Amplivox private release**

Interacoustics AD229 (same interface as AD629)

**6/5/12 – Version 4.6.0.12 –private release**

Fixed AD229 distortion and electrical control logic error

**6/12/12 – Version 4.6.0.13 – Public release**

Otovation Otosphere – Free Field test, cal date format and EEPROM slot table debugging

**7/10/12 – Version 4.6.0.14 – Private release**

Otovation Otosphere – turn off 1 kHz reference tone after Speech Level dialog closes

**7/30/12 – Version 4.6.0.17 – Private release**

Consolidated IA interfaces to IAMain\_IO.dll  
IA Calibration DLL pack update  
Added Grason Stadler Audiostar Pro interface

**8/8/12 – Version 4.6.0.19 – Private release**

Otovation Otosphere – setup logic for free field speaker angles

**8/9/12 – Version 4.6.0.20 – Private release to Otovation**

DD45 support for SIB Primus

**8/20/12 – Version 4.6.0.13a – Fix for White Noise using HF Phones – Public Release**

**8/20/12 – Version 4.6.0.21 – Private release**

Added TDH50 support for AudiostarPro

**8/30/12 – Version 4.6.0.22 – Private release**

Added Notes field to Otosphere setup dialog  
Added Cal Data left and right to Otosphere setup dialog  
Stored WN cal to DI and US signals

**9/19/12 – Version 4.6.0.23 – Private release**

Revised logic for AudiostarPro impedance and extended transducer types

**10/3/2012 – Version 4.6.0.24 – Private release**

Changes for Otosphere error ER18 and ER20 handling

**10/19/12 – Version 4.6.0.25 – Private release**

Added new “Validation Margins” feature

**10/19/12 – Version 4.6.0.26 – Private release**

Interacoustics AC40 interface support  
GN Otometrics Astera interface support  
GN Otometrics Aurical Aud interface support

**11/14/12 – Version 4.6.0.27 – Private release**

- Validates the Audiostar Pro coupler dependent transducer setting against the corresponding coupler selection in the equipment setup. For example, if the Audiostar is set to EAR3A artificial ear, the coupler setting the the ACS equipment setup must match (IEC-711) or an error message will be shown.

- The dialog box prompting the technician to select the transducer type to be used for electrical tests will default to headphones and automatically timeout after 60 seconds.

- Set the next calibration date on the audiometer to current date plus 12 months. This logic applies to all IA models.

- HF Phones HDA-200 RETSPL value set as follows:

Case 1: 18K = 83, 20K = 105

GSI-61

Audiostar Pro

AC40

AD629

AD229

And new IA audiometers

Case 2: 18K = 72, 20K = 86

All other IA audiometers

GN Otometrics (Aster/Aurical) Logic changes to fix shutdown crash

SIB\_PRIMUS

- Post Calibration Actions.

- Copy calibration data to EEPROM on transducer plug. A new set of control buttons have been added for this under the "Setup Audiometer" dialog box in the Audiometer setup panel.

#### **12/15/12 – Version 4.6.0.28 – Private release**

IAMain\_IO.dll – fixed transducer indexing for Callisto (Inserts could not be selected)

New IA Calibration Pack from Peter Garup (Audiostar Pro NBN changes)

IAMain\_IO.dll v1.0.0.7

Titan\_IO.dll v1.0.0.13

#### **1/22/13 – Version 4.6.0.29 – Private release**

Added interface support for Interacoustics AD226

Interacoustics Callisto – added support for HAD-200

Real\_Ear Primus – detection of transducer logic and support for PI-1

GNO Astera/Aurical – fixed disconnect/reconnect issue

IAMain\_IO.Dll version 1.0.0.8

Titan\_IO.dll version 1.0.0.14

#### **4/5/13 – Version 4.6.0.30 - Public Release**



Fixed EAR5A transducer selection error on some devices  
Lock Audiostar Pro keyboard when in calibration mode  
Set Callisto HF Retspl for 18K and 20K same as Audiostar Pro  
Fixed issue with Adjustment threshold and cal value/offset (resulted in the calibration value not being adjusted (pre = post) under certain boundary conditions).  
Added logic to only list transducer information on certificate if one or more tests are selected in the certificate setup options.  
Added impedance on certificate (just to the right of the firmware version).  
Changed the test order for fully automatic Tymp (moved volume and pressure tests to end) to allow uninterrupted calibration of probe and ipsi levels.  
Added 2.5 second delay after saving cal data on MA-41/42 before closing connection  
IAMain\_IO.Dll version 1.0.0.9

#### **4/20/13 – Version 4.6.0.31 - Public Release**

Added interface support for Inventis Models (Bell/Harp/Piano/Piccolo)  
Faster performance when “Search for Audiometer” used (SQL query)  
AudiostarPro – refined logic for determining Impedance value on certificate  
Added selection for Speech standard in Test Configuration  
Selectable Speech standards for IEC/ANSI/Norway/Sweden)  
Speech standard shown on certificate  
View Calibration – Edit button allows standards to be selected before printing certificate  
Narrowband 1/3 Octave switch added to test Configuration  
White Noise bandwidth switch added to test Configuration  
Equip setup - added option for insert mic to be connected to 2nd mic channel (ACS-100 only)  
Otovation Otosphere – Slot table tool shows values in red if MaxHL exceeded  
Otovation Otosphere – sound card playback volume set to 100 during Speech Level cal  
Otovation Otosphere – send ANSI command after connecting to unit  
Signal Analyzer – added spectrum and time series plots  
Signal Analyzer – added “Minimize” button to shrink display

#### **8/16/13 – Version 4.6.0.32 - Private Release**

Allow Tymp Reflex tests with no Contraphone  
Revised TENnoise RETSPL values for AudiostarPro  
Astera/Aurical – additional error checking for calibration values

#### **9/23/13 – Version 4.6.0.32A - Public Release**

Revised DD45 RETSPL values per updated values from WDH  
Revised HDA200 RETSPL values at 18 and 20K for Affinity/Equinox per WDH requirements

#### **10/15/13 – Version 4.6.0.32B - Public Release**

Added serial interface support for Benson NEXT audiometer

#### **10/26/13 – Version 4.6.0.32C - Public Release**

Developmental testing changes for Amplivox models

#### **11/08/13 – Version 4.6.0.32D - Public Release**

Corrected linearity validation logic issue

**11/13/13 – Version 4.6.0.32E - Public Release**

Added support for the Sennheiser HDA-300 HF phones (not yet available for the Interacoustic models when under automatic control).

Added option to include individual calibration information in the Select Audiometer results listing.

Added clipboard support for the contents of the Select Audiometer and Search Audiometer results listing; allows the search results to be copied and pasted into Excel or other program.

Added user selectable path for the microphone and mastoid files. This allows the .mic and .mst files to be located in a different folder than the main program location. This new setting is in the Equipment Setup dialog box near the bottom of the panel.

Added logic to validate the microphone and mastoid calibration dates and warn the user if any are expired or due for calibration within 30 days. To accommodate this, the microphone and mastoid calibration date field has been changed to a calendar control - you will be prompted to convert the old date string to a calendar date code when editing an existing .mic or .mst file.

Added Run/Stop toolbar buttons to the Spectrum and Time Series plots in the Signal Analyzer. Allows you to pause the analyzer without changing back to the main analyzer control tab.

**2/27/14 – Version 4.6.0.33 - Public Release**

Fixed problem with mic calibration when option selected to allow insert mic to be connected to 2nd mic channel (ACS-100 only).

Otovation Otosphere - added slight delay after starting NBN/WN/SN before measuring level to allow signal to fully start.

Improved Pulse Width measurement logic to better deal with cases when the actual frequency differs significantly from the test frequency.

**3/12/14 – Version 4.6.0.34 - Public Release**

Added support for the new transducers (IP30/HDA300/B81) in the Audiostar Pro  
Transducer Impedance detection and notation for the Audiostar Pro  
Functionality from Titan\_IO.dll merged into IAMain\_IO.dll. Titan\_IO.dll no longer used.

**4/22/14 – Version 4.6.0.35B – Private test release**

Fixed problem with causing negative instrument serial numbers to be displayed.

Added new Free Field Retspl values for 2010 spec

Added option to include individual calibration information in the Audiometer Search results listing.

Fixed missing headers on certificate in certain scenarios

Added support for Siemens Unity 3

**5/7/14 – Version 4.6.0.36 - Public release**

Updated Retspl values for DD45 when using a 318-1 coupler (artificial ear) per Interacoustics tables

**5/15/14 – Version 4.6.0.36A - Public release**

Fixed audiometer label issue on certificate

**5/20/14 – Version 4.6.0.36B - Public release**

Changes for all Amplivox instruments

Changes for Tremetrics RA-800

Fix for Real Ear Primus and Unity 3 Electrical test logic

Fix for Click measurement

**7/17/14 – Version 4.6.0.36G - Private release**

Added slight delay on signal analyzer start when click option selected to avoid startup transients

**7/29/14 – Version 4.6.0.37 - Public release**

Added DD45 and IP30 to list of selectable Contra phones

Amplivox PC850 and SD270 models listed separately

Added Holmco 8103 interoctave RETSPL values and extended Speech types

**9/3/14 – Version 4.6.0.37A - Private release**

Added support for Interacoustics AT235 (V3)

**9/29/14 – Version 4.6.0.37B - Private release**

Updates for Interacoustics AT235 (V3)

**10/9/14 – Version 4.6.0.37C - Private release**

Updated interface components to fix compatibility problem with dispenser software for Interacoustics Equinox, Affinity and Calisto.

**10/21/14 – Version 4.6.0.38 – Public release**

Implemented new Cavity Volume Calibration logic for AT235 V3

Added Pressure Calibration buttons (0 and -200)

Added Contra Phone transducer setup capability

**11/8/14 – Version 4.6.0.38A – Private release**

Changes for AT235 V3 tympanometry tests

**11/12/14 – Version 4.6.0.38B – Private release**

Changes for AT235 V3 tympanometry tests

**11/14/14 – Version 4.6.0.38C – Private release**

Added Smart Diagnostic Devices – Smart Tone audiometer (manual calibration)  
Removed automatic control for GN Astera and Aurical (firmware issues)  
Changed default ACS Logo

#### **11/26/14 – Version 4.6.0.39 – Public release**

Added support for Tymp Probe transducer setup for Interacoustics models  
Added individual Test HL settings for special speech noise types  
Reflex LBN/BBN/HBN noise bandwidth set to 100 to 10K  
Fixed Validation Margin logic issue with Reflex test  
Narrowband spec added to specifications list on Certificate  
Removed warning message about Contra phone setup when only Probe is enabled  
White Noise bandwidth test options now (250-5K or 100-10K)

#### **1/6/15 – Version 4.6.0.39A – Private release**

User Preferences – Test Options: added option “Allow Editing Configuration” to prevent inadvertent cfg changes in a production environment. If enabled, Config filename is stored in the calibration notes.  
Interacoustics Factory Specification for Tympanometry added  
Interacoustics – special buttons “Print ANSI” and “Print IEC” added to View Cal – Specs tab  
Otovation – added support for DD65 transducers  
Update for Inventis firmware v6.00

#### **1/21/15 – Version 4.6.0.39B – Private release**

Changes for AuditData models  
Primus Fitting Unit +  
ILTASS support  
Write Script capability

#### **2/3/15 – Version 4.6.0.39C – Private release**

Support for AS608, MA-25 (V2), MA-27 (V2), Colson K20  
Changes for AuditData instrument setup utility

#### **3/15/15 – Version 4.6.0.39H – Private release**

Fixed problem causing excessive distortion at 6kHz and above with large microphone corrections

#### **4/14/15 – Version 4.6.0.40 – Public release**

Added automatic support for the following instruments:  
Maico MA-25  
Maico MA-51  
Maico MA-52  
Interacoustics AS608  
Colson K20

Siemens Unity 3 – Bone calibration written to both left and right channels

#### **5/1/15 – Version 4.6.0.41 – Public release**

Support for GSI Tymptest Pro and Maico Touch Tymptest  
IA Cal Pack (Apr 29 2015), IA\_Main v1.0.0.22  
Update for DT API interface libraries

**5/8/15 – Version 4.6.0.41A – Private release**

Support for Contra phone on Probe Box or Device for TymptestPro/Touch Tymptest/AT235  
IA Cal Pack (Apr 29 2015), IA\_Main v1.0.0.26  
Siemens Unity 3 – Bone PI and MO calibration values stored in both left and right  
Siemens Unity 3 – logic change to turn signal back on after adjustment warning dialog box shown.  
Siemens Unity 3 – fix label on certificate for insert phones

**6/12/15 – Version 4.6.0.41C – Private release**

TouchTymptest - Fixed problem reading transducer info after change/reboot.  
Tymptest Pro - Read/Write probe serial number  
Click Level - set level equal to tone peak to peak level  
Probe output level - max/min cal values set to +/-15

**6/17/15 – Version 4.6.0.41D – Private release**

Click Test – removed validation of click target level

**6/18/15 – Version 4.6.0.41E – Private release**

Siemens Unity 3 – Bone Osc cal date set for both left and right cal tables.

**6/24/15 – Version 4.6.0.41F – Private release**

Siemens Unity 3 – Bone Osc Retspl set for both left and right tables.  
Inventis Models – Turn pulse mode off when exiting pulse width test  
Validation – fixed problem validating tymptest reflex frequency

**7/10/15 – Version 4.6.0.42 – Public release**

**9/30/15 – Version 4.6.0.42 – Removed due to issues with Unity 2 distortion – update will be available soon.**

Support for Audit Data Primus Pro  
Interacoustics Eris platform instruments  
ABR and OAE tests  
IA calibration pack update – fixes Unity 2/Affinity/Equinox extended range error (distortion issue)

**2/12/16 – Version 4.6.0.43G – Private Release (IA Main v1.0.0.29)**

Support for GSI Tymptest  
Support for GSI Pello  
Support for Radioear DD450 HF phones  
Changed name of Maico easyScreen to screenEar  
Support for IEC 60645 2012  
Support for IOWA probe model

**3/28/16 – Version 4.6.0.44 – Private Release (IA Main v1.0.0.31)**

Fix for problem with distortion level for Unity 2/Equinox/Affinity  
Support for Interacoustics AA222 (Model 1078)  
Corrected Tymp Compliance spec limits displayed on View calibration page  
Corrected Welch Allyn Audioscope retspl values  
Changed Maico to MAICO

**6/9/16 – Version 4.6.0.45 – Public Release (IA Main v1.0.0.33)**

Corrected message box text when failing to connect to certain models to require VC++ 2012 runtime instead of VC++ 2008.

**7/13/16 – Version 4.6.0.45B – Public Release (IA Main v1.0.0.33)**